



BSI Standards Publication

Foodstuffs — Determination of lipophilic algal toxins (okadaic acid group toxins, yessotoxins, azaspiracids, pectenotoxins) in shellfish and shellfish products by LC-MS/MS

National foreword

This British Standard is the UK implementation of EN 16204:2012.

The UK participation in its preparation was entrusted to Technical Committee AW/275, Food analysis - Horizontal methods.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2012. Published by BSI Standards Limited 2012

ISBN 978 0 580 72863 1

ICS 67.120.30

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 May 2012.

Amendments issued since publication

Date	Text affected
------	---------------

Licensed copy. University of Manchester Library. The University of Manchester Library. Version control as of 06/05/2016

EUROPEAN STANDARD

EN 16204

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2012

ICS 67.120.30

English Version

Foodstuffs - Determination of lipophilic algal toxins (okadaic acid group toxins, yessotoxins, azaspiracids, pectenotoxins) in shellfish and shellfish products by LC-MS/MS

Produits alimentaires - Dosage des toxines algales lipophiles (toxines du groupe acide okadaïque, yessotoxines, azaspiracides, pectenotoxines) dans les coquillages et les produits à base de coquillages par CL-SM/SM

Lebensmittel - Bestimmung der lipophilen Algentoxine (Okadasäuregruppen-Toxine, Yessotoxine, Azaspirosäuren, Pectenotoxine) in Schalentieren und Schalentiererzeugnissen mit LC-MS/MS

This European Standard was approved by CEN on 20 April 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Principle	5
4 Reagents	5
5 Apparatus	7
6 Procedure	8
6.1 Preparation of samples	8
6.1.1 General	8
6.1.2 Raw samples	8
6.1.3 Cooked samples	8
6.2 Homogenization and extraction	8
6.3 Hydrolysis	8
7 HPLC-MS/MS analysis	9
7.1 General	9
7.2 HPLC operating conditions (chromatography under acidic conditions)	9
7.3 HPLC operating conditions (chromatography under basic conditions)	10
7.4 Mass spectrometric operating conditions	10
7.5 Calibration curve	10
7.6 Determination of algal toxins in sample test solutions	11
7.7 Quality control measures for sequences	11
8 Calculation	12
8.1 Peak identification	12
8.2 Quantitative determination by means of external calibration and matrix correction	12
8.3 Description of matrix correction	13
8.4 Calculation of the total toxicity	14
9 Precision	14
10 Test report	14
Annex A (informative) Precision data	15
A.1 Details on the inter-laboratory study	15
A.2 Recovery	28
Annex B (informative) Examples for suitable MS detection conditions	29
B.1 Examples suitable for SCIEX API 4000 or API 4000 Q-Trap	29
B.2 Examples suitable for Waters (Micromass) TSQ Ultima	31
B.3 Examples suitable for Thermo Fisher TSQ Quantum Ultra	33
B.4 Examples suitable for Agilent 6410 or 6460 QQQ	35
Annex C (informative) Typical chromatogram	37
Bibliography	38

Foreword

This document (EN 16204:2012) has been prepared by Technical Committee CEN/TC 275 "Food Analysis - Horizontal methods", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2012, and conflicting national standards shall be withdrawn at the latest by November 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

Lipophilic marine biotoxins are the most frequently occurring algal toxins in Europe and are produced by certain marine dinoflagellates. They can accumulate in filter-feeding bivalves reaching highly toxic levels and, after consumption, may cause harm in humans such as nausea, vomiting and diarrhoea.

Commission Regulation 15/2011 stipulates LC-MS/MS as the reference methodology and refers to a method validated by the EU-RL Network. The method presented in EN 16204 is proposed as an alternative method to the one validated by EU-RL.

1 Scope

This European Standard specifies a multi-reference method for the determination of lipophilic algal toxins (fat-soluble algal toxins produced by some dinoflagellates) in raw shellfish and shellfish products including cooked shellfish, by liquid chromatography coupled to tandem mass spectrometry LC-MS/MS [1], [2], [3]. This method has been validated in an inter-laboratory study consisting of three parts via the analysis of both naturally contaminated homogenates of blue mussel and spiked extracts of blue mussel, oyster and clam. For further information on the validation, see Annex A. Additional studies have investigated further matrices (see [4], [5]).

The detection limit for toxins of the okadaic acid group, azaspiracids and pectenotoxins was determined to be 6 µg/kg shellfish meat and for yessotoxins 10 µg/kg shellfish meat.

Quantitative determination of okadaic acid (OA), pectenotoxin-2 (PTX-2), azaspiracid-1 (AZA-1) and yessotoxin (YTX) can be carried out directly by means of standard substances available commercially. Assuming an equal response factor, okadaic acid is used for the indirect quantitative determination of the two dinophysistoxins dinophysistoxin-1 (DTX-1) and dinophysistoxin-2 (DTX-2); likewise azaspiracid-1 (AZA-1) is used for the indirect quantitative determination of azaspiracid-2 (AZA-2) and azaspiracid-3 (AZA-3), while YTX is used for homo-yessotoxin, 45-OH-yessotoxin and 45-OH-homo-yessotoxin, and PTX-2 for pectenotoxin-1 (PTX-1).

The limit of quantification (LOQ) for toxins of the okadaic acid group, azaspiracids and pectenotoxins was determined to be 20 µg/kg shellfish meat and for yessotoxins 35 µg/kg shellfish meat.

By means of hydrolysis [6], the esters of okadaic acid, DTX-1 and DTX-2 can also be determined quantitatively as the corresponding free acids.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696:1995, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

3 Principle

Remove the shellfish meat from the shell and homogenize the total shellfish meat. Extraction is carried out with aqueous methanol ($\varphi = 80\%$). Separation is performed on a HPLC reverse-phase column provided with a binary gradient and detection is carried out by means of tandem mass spectrometry using triple quadrupole technology. The concentration of lipophilic toxins is determined by means of external calibration.

4 Reagents

If not otherwise specified, reagents of analytical grade and solvents suitable for LC-MS/MS shall be used. Water shall be distilled in glass vessels or demineralised before use, or shall be of equivalent purity according to EN ISO 3696:1995. Since the use of this method involves reagents harmful to health, appropriate precautionary and protective measures such as avoiding skin contact and using an extractor hood shall be taken.

4.1 Aqueous methanol ($\varphi = 80\%$).

NOTE The validation data of this method have been elaborated with 80 % aqueous methanol. However, it has been shown (see [4], [5]) that equivalent results can be obtained when using 100 % methanol.

4.2 Acetonitrile

4.3 Sodium hydroxide, $c(\text{NaOH}) = 2,5 \text{ mol/l}$.

4.4 Hydrochloric acid, $c(\text{HCl}) = 2,5 \text{ mol/l}$.

4.5 Formic acid, 98 % to 100 % w/w.

4.6 Ammonium formate

4.7 Nitrogen, gaseous, min purity: 5,0.

4.8 Ammonium hydrogen carbonate

4.9 HPLC mobile phase 1 (chromatography under acidic conditions)

4.9.1 Eluent A1

Dissolve 126 mg (to give a 2 mmol/l solution) of ammonium formate (4.6) and 2 ml (to give a 50 mmol/l solution) of formic acid (4.5) in 50 ml of water and fill up to 1 000 ml with water. If necessary, filter the eluent using a 0,45 μm membrane filter.

4.9.2 Eluent B1

Dissolve 126 mg (to give a 2 mmol/l solution) ammonium formate (4.6) and 2 ml (to give a 50 mmol/l solution) of formic acid (4.5) in 50 ml of water. Add 950 ml of acetonitrile (4.2) and filter the eluent using a 0,45 μm membrane filter, if required.

4.10 HPLC mobile phase 2 (chromatography under basic conditions)

4.10.1 Eluent A2

Dissolve 395 mg (to give a 5 mmol/l solution) of ammonium hydrogen carbonate (4.8) in 1 000 ml of water. If necessary, filter the eluent using a 0,45 μm membrane filter.

4.10.2 Eluent B2

Dissolve 395 mg (to give a 5 mmol/l solution) of ammonium hydrogen carbonate (4.8) in 50 ml water. Add 950 ml of acetonitrile (4.2) in portions of about 100 ml. Shake vigorously after each portion added. If necessary, filter the eluent using a 0,45 μm membrane filter.

4.11 Toxin-free shellfish homogenate

To estimate and determine matrix effects, standard solutions in matrix are prepared. The shellfish homogenate required for this purpose is prepared from shellfishes that have been proved free from toxins.

4.12 Reference substances¹⁾

NOTE During the validation study, the following reference substances were available. If other certified reference substances should become available in the future, the use of these substances is recommended.

4.12.1 Okadaic acid (OA)

4.12.2 Pectenotoxin-2 (PTX-2)

1) Reference substances can be purchased from National Research Council Canada (NRC), Institute for Marine Bioscience, Halifax, for instance (<http://www.nrc-cnrc.gc.ca>). This is an example for suitable products available commercially. This information is given for the convenience of users of this standard and does not constitute an endorsement by CEN of these products.

4.12.3 Azaspiracid-1 (AZA-1)

4.12.4 Yessotoxin (YTX)

4.12.5 Multi-toxin standard solutions

Prepare standard calibration solutions (multi-toxin standard) in aqueous methanol (4.1). The following concentrations are recommended:

- OA, AZA-1, PTX-2: 1,5 ng/ml; 2,5 ng/ml; 5,0 ng/ml; 10,0 ng/ml; 15,0 ng/ml and 25,0 ng/ml;
- YTX: 3,0 ng/ml; 5,0 ng/ml; 10,0 ng/ml; 20,0 ng/ml; 30,0 ng/ml; and 50,0 ng/ml.

4.12.6 Multi-toxin standard solutions in matrix

Prepare a solution with a standard concentration level by dilution in a toxin-free methanolic shellfish extract (6.2) to estimate or determine matrix effects. The recommended concentration levels are:

- OA, AZA-1, PTX-2: 15,0 ng/ml;
- YTX: 30,0 ng/ml.

5 Apparatus

Usual laboratory glassware and equipment and, in particular, the following:

- 5.1 Mechanical high-speed blender or homogenizer** (e.g. Grindomix, Ultra-Turrax®²⁾.
 - 5.2 Centrifuge** (working at minimum 2 000 g) **and centrifuge tubes** (volume 20 ml).
 - 5.3 Shaker** (e.g. Vortex).
 - 5.4 Heat block.**
 - 5.5 Analytical balance**, accuracy to the nearest 0,1 mg.
 - 5.6 Volumetric flask**, 20 ml or 10 ml.
 - 5.7 Graduated cylinder and suitable pipettes.**
 - 5.8 High Performance Liquid Chromatography (HPLC) system**, capable of gradient elution.
 - 5.9 HPLC vials.**
 - 5.10 Triple-Quadrupol-LC-MS/MS system.**
 - 5.11 Analytical Reversed Phase Column**, e.g. RP C18, particle size 3 µm, 150 mm (length) × 2 mm (diameter) or C8, 50 mm (length) × 2 mm (diameter), 3 µm particle size, (only for acidic conditions).
- Optionally, an appropriate guard column may be used.
- 5.12 Syringe or membrane filter** (e.g. regenerated cellulose membranes with a pore diameter of 0,45 µm).

2) Ultra Turrax® is an example for a suitable product available commercially from various suppliers. This information is given for the convenience of users of this standard and does not constitute an endorsement by CEN of this product.

5.13 Syringe for filter system, e.g. 1 ml.

5.14 High-speed table centrifuge (up to $10\,000 \times g$).

6 Procedure

6.1 Preparation of samples

6.1.1 General

Storage of shellfish samples in frozen state (at -18 °C) for up to a year has no negative influence (see [7]) on the results obtained with this method.

6.1.2 Raw samples

After receipt, the raw samples should be shelled and drained, and then frozen until used for analysis or shall be extracted immediately.

6.1.3 Cooked samples

The sample is cooked/steamed after receipt then drained and shelled. If it is not immediately treated, it can be cooled down until 24 h.

To cook the shellfish, heat a sufficient amount of water (approximately 2 l to 3 l water per 1 kg of shellfish) to the boiling point. When the water is boiling, place shellfish into the water and cook under further addition of heat for approximately 3 min. After cooking, remove the shellfish meat from the open shells and homogenize.

6.2 Homogenization and extraction

For the determination of lipophilic algal toxins, homogenize 100 g to 200 g shellfish meat or the meat of 50 to 100 whole shellfish in order to ensure representative data. If only small amounts are available, the whole sample amount is used.

Homogenize shellfish meat by means of a mechanical blender (5.1) and prepare immediately or store at -18 °C .

Weigh $(2,00 \pm 0,02)$ g of homogenate into a centrifuge tube (5.2). Add 9 ml of aqueous methanol (4.1) and mix for 1 min using a blender (5.1) at $13\,500\text{ min}^{-1}$. Subsequently, centrifuge (5.2) the mixture at approximately $2\,000 \times g$ at room temperature for 5 min. Decant the supernatants into a 20 ml volumetric flask. Repeat the extraction. Decant the second supernatants into the 20 ml volumetric flask and dilute to the mark with aqueous methanol (4.1). Subsequently, filter the extract using a $0,45\text{ }\mu\text{m}$ membrane filter (5.12). As an alternative to filtration, a high-speed table centrifuge with $10\,000 \times g$ may be used for the same purpose (5.14).

If only a small sample amount is available, $(1,00 \pm 0,01)$ g homogenate may be extracted twice with 4,5 ml of methanol (4.1) and made up to 10 ml.

6.3 Hydrolysis

The hydrolysis step is necessary for determination of toxins of the okadaic acid group, bound as esters. From this step, the sum of free and bound toxins of the okadaic acid group is determined. The bound toxins are determined as the difference of the sum (after hydrolysis) and the free toxins (without hydrolysis).

Pipette 1,0 ml of methanolic extract (6.2) into a vial (5.9). Add 125 μl of sodium hydroxide (4.3) and heat in a heat block (5.4) at $(76 \pm 2)\text{ °C}$ for 40 min. After cooling the solution to room temperature, add 125 μl of

hydrochloric acid (4.4) for neutralization and shake for 30 s (5.3). Subsequently, filter the samples through a 0,45 µm membrane (5.12).

Other amounts can also be used for hydrolysis. Care should be taken to ensure following ratios:

— Crude/raw extract / NaOH (2,5 mol/l) / HCl (2,5 mol/l) corresponding to 1 / 0,125 / 0,125.

During hydrolysis vials shall be closed firmly (boiling point of methanol is 65 °C). The loss of methanol during hydrolysis can be checked by weighing the vials before and after heating.

7 HPLC-MS/MS analysis

7.1 General

The measuring conditions stated are merely indicative; they may be adjusted to the respective local conditions. Triple Quadrupole Tandem Mass Spectrometry should be used. Analytes that cannot be distinguished by mass spectrometry (e.g. OA and DTX-2) shall be baseline separated by means of chromatography.

Before starting measurements, the HPLC system should be thoroughly rinsed (use starting conditions). Moreover, it is recommended to perform a test run for at least one sample before starting the sequence.

7.2 HPLC operating conditions 1 (chromatography under acidic conditions)

Chromatographic conditions may be chosen freely. The acceptable minimum retention time (for the analyte under examination) is twice the retention time for the void volume of the column.

EXAMPLE 1 When using the C18-column specified in 5.11 and the mobile phases A1 (4.9.1) and B1 (4.9.2), the conditions in Table 1 were found to be appropriate:

Table 1 — HPLC operating conditions 1

Time min	Flow rate ml/min	Mobile phase A1 %	Mobile phase B1 %
0	0,2	60	40
6,0	0,2	10	90
14,5	0,2	10	90
15,5	0,2	60	40
21,0	0,2	60	40

- Flow rate mobile phase (column): 0,2 ml/min
- Injection volume: 5 µl to 10 µl
- Column oven temperature (including the guard column): 40 °C

EXAMPLE 2 When using the C8 column, specified in 5.11, and the mobile phases A1 (4.9.1) and B1 (4.9.2), the conditions in Table 2 were found to be appropriate:

Table 2 — HPLC operating conditions 2

Time min	Flow rate ml/min	Mobile phase A1 %	Mobile phase B1 %
0	0,2	70	30
8,0	0,2	10	90
11,0	0,2	10	90
11,5	0,2	70	30
16,0	0,2	70	30

- Flow rate mobile phase (column): 0,2 ml/min
- Injection volume: 5 µl to 10 µl
- Column oven temperature (including the guard column): 40 °C

7.3 HPLC operating conditions 2 (chromatography under basic conditions)

When using the column specified in 5.11 and the mobile phases A2 (4.10.1) and B2 (4.10.2), the conditions in Table 3 were found to be appropriate (see also Figure C.1):

Table 3 — HPLC operating conditions 3

Time min	Flow rate ml/min	Mobile phase A2 %	Mobile phase B2 %
0	0,2	80	20
15,0	0,2	10	90
20,0	0,2	10	90
21,0	0,2	80	20
26,0	0,2	80	20

- Flow rate mobile phase (column): 0,2 ml/min
- Injection volume: 5 µl to 10 µl
- Column oven temperature (including the guard column): 40 °C

7.4 Mass spectrometric operating conditions

The measurement may be performed using various instruments and instrument parameters. Some instrument parameters are listed in Annex B. These conditions have been shown to provide satisfactory results.

7.5 Calibration curve

Prepare a calibration curve which includes at least 6 concentration levels (excluding the origin) and which covers a concentration range from e.g. 1,5 ng/ml to 25 ng/ml (or 3,0 ng/ml to 50 ng/ml for YTX). If a larger

concentration range is required, the calibration line may be sectioned for several ranges or a weighted calibration may be carried out.

The lowest standard should be at or equal to the LOQ.

Samples more highly concentrated than the highest calibration level can also be diluted with blank shellfish extract to obtain the calibration range.

The ratio between intercept and slope is calculated and used as a quality parameter. The absolute value of this ratio corresponds to the concentration plotted on the x-axis and shall not be greater than half the quantification limit. Additionally, the correlation coefficient (r^2) should be equal to or greater than 0,985.

7.6 Determination of algal toxins in sample test solutions

Inject aliquots of the sample test solution (6.2), (6.3) into the HPLC system in an appropriate sequence.

7.7 Quality control measures for sequences

All samples subject to quantitative determination of lipophilic algal toxins shall be analyzed at least by duplicate injection. The repeatability limit of the duplicate injection shall be smaller than $2,8 \times r$, where, r stands for the repeatability standard deviation of the extract.

To check the quality of chemical analysis, it has proven necessary to run simultaneously at least one QM sample per sequence. This sample should contain a known concentration of lipophilic algal toxins such as a sample with assigned value used in an inter-laboratory trial, a certified reference material (e.g. NRC Mus-b), a laboratory reference material, or a blank extract/blank homogenate spiked with standards. When tolerance or warning limits are defined, the different repeatability standard deviations of extracts and homogenates shall be taken into account.

The matrix effects expected within a sequence shall be corrected as described in 8.3.

For the determination of lipophilic algal toxins the adoption of a defined sequence has been proven suitable.

The following operations layout is recommended:

a) For sequences including more than 20 samples or exceeding a run time of 12 h:

- 1) calibration standards;
- 2) matrix correction standard;
- 3) QM sample;
- 4) samples ($n > 20$, first injection);
- 5) calibration standards;
- 6) matrix correction standard;
- 7) QM sample;
- 8) samples ($n > 20$, second injection);
- 9) optionally: one calibration standard.

b) For sequences including less than 20 samples or a run time below 12 h:

- 1) samples ($n < 20$, first injection);

- 2) calibration standards;
- 3) matrix correction standard;
- 4) QM sample;
- 5) samples (n < 20, second injection);
- 6) optionally: one calibration standard.

Bracketing of standards or repeat injection of standards in between samples is also allowed for drift correction, although this has not been carried out in the validation study.

8 Calculation

8.1 Peak identification

The qualitative detection of lipophilic toxins is carried out by comparison of the retention time of the analytes in the sample with those of the standard substances or reference material. A confirmatory transition may be acquired for definitive identification of the analyte.

8.2 Quantitative determination by means of external calibration and matrix correction

The quantitative determination is performed according to the method of the external standard by integration of the peak areas in relation to the calibration line of the standard solutions.

Assuming an equal response factor, okadaic acid is used for the quantitative determination of the dinophysistoxins DTX-1 and DTX-2; likewise AZA-1 is used for the quantitative determination of AZA-2 and AZA-3. YTX is used for homo-yessotoxin, 45-OH-yessotoxin and 45-OH-homo-yessotoxin, and PTX-2 for PTX-1, if applicable. For the determination of the esters of OA, DTX-1 or DTX-2, quantification of the free acids is carried out before and after hydrolysis. It would be appropriate to use certified standards for the other analogues as soon as they are commercially available.

Evaluation is based on the linear equation of the regression line of the individual lipophilic algal toxins. The correction of matrix effects is performed with the aid of a toxin-free shellfish extract of the same matrix strength which is spiked with standard substances (4.12.6).

Calculate the mass concentration of the individual toxins in a sample using Formula (1):

$$c = \frac{(A - b)}{a} \times F \times \text{Matrix correction} \quad (1)$$

where

- c is the mass concentration of the analyte in micrograms toxin per kilogram of sample;
- A is the peak area, in units;
- b is the Y-intercept of the regression line;
- a is the slope of the regression line;
- F is the dilution factor:
 - 10 for measurement of the crude extract (6.2) without hydrolysis;
 - 12,5 for measurement of the crude extract (6.3) with hydrolysis.

With the hydrolysis step, the sum of free and bound toxins of the okadaic acid group is determined. The bound toxins are determined as the difference of the sum (after hydrolysis) and the free toxins (without hydrolysis).

8.3 Description of matrix correction

The validation study has clearly outlined that matrix correction with a single species can be applied to other species (see [8]) with a single matrix matched concentration level.

Other studies have shown that full matrix matched calibration may also be used to correct for matrix effects.

Where single point matrix correction is applied, the prepared matrix standard (see 4.12.6) is analyzed. The concentration of the matrix standard is determined by means of the regression line. The concentration found is an indication of the apparent recovery (matrix effect) of the analytes and reflects the response suppression or enhancement of the analytes in the matrix. The concentration determined in the unknown sample is corrected with the apparent recovery.

EXAMPLE 1 AZA-1 in the matrix standard:

$$c[\text{matrix AZA1}] = \frac{(A - b)}{a} \quad (\text{external calibration}) \quad (2)$$

$$c[\text{matrix AZA1}] = 7,5 \text{ ng/ml} \quad 100 \% \text{ recovery (spiked): here: 15 ng/ml} \quad (3)$$

$$\text{apparent recovery [matrix AZA-1]} = 50 \% \quad (4)$$

If, for instance, the apparent recovery of AZA-1 in matrix is determined to be 50 %, the AZA-1 concentration of each unknown sample shall be corrected by applying the factor for correction of apparent recovery.

EXAMPLE 2 AZA-1 matrix correction in the unknown sample:

$$c [\text{AZA1 corr}] = \frac{c \text{ AZA1 (external calibration)} \times 100}{\text{apparent recovery (here : 50)}} \quad (5)$$

Matrix effects are determined for the individual toxins as follows:

- OA: Recovery of the 15 ng/ml standard in matrix; the OA can also be corrected using a dilution (e.g. 1/20 in matrix) of Mus-b³⁾ with blank shellfish extract;
- DTX-2: Assumption of the same matrix effect as for OA;
- DTX-1: Recovery of DTX-1 in NRC Mus-b³⁾ (e.g. 1/20 in matrix) or assumption of the same matrix effect as for OA;
- PTX-2: Recovery of the 15 ng/ml standard in matrix;
- PTX-1: Assumption of the same matrix effect as for PTX-2;
- YTX: Recovery of the 30 ng/ml standard in matrix;
- 45-OH-YTX, homo-YTX and 45-OH-homo-YTX: Assumption of the same matrix effects as for YTX;

3) The certified reference material Mus-b is obtained from a mix of mussels digestive gland and algal cells.

- AZA-1: Recovery of the 15 ng/ml standard in matrix;
- AZA-2 and AZA-3: Assumption of the same matrix effect as for AZA-1.

It would be appropriate to use certified standards for the other analogues as soon as they are commercially available.

8.4 Calculation of the total toxicity

Concentrations of determined toxins shall be converted to the toxic equivalence of the key toxin of that particular toxin group by applying the toxic equivalence factors (TEF) proposed by the EFSA (see [9], [10], [11], [12]).

9 Precision

Details of an inter-laboratory test on the precision of the method according to ISO 5725-2, -4 and -6 (see [13], [14], [15]) are given in Annex A. The values derived from this inter-laboratory test may not be applicable to concentration ranges and/or matrices other than those given in Annex A.

10 Test report

The test report shall contain the following data:

- a) all information necessary for the identification of the sample (kind of sample, origin of sample, designation);
- b) reference to this European Standard;
- c) date and type of sampling procedure (if known);
- d) date of receipt;
- e) date of test;
- f) test results and the units in which they have been expressed;
- g) any particular points observed in the course of the test;
- h) any operations not specified in the method or regarded as optional, which might have effected the results.

Annex A (informative)

Precision data

A.1 Details on the inter-laboratory study

The data given in Table A.1 to Table A.12 were obtained in an international inter-laboratory study (see [8], [16]) organized by the Working Group “Phycotoxins” of the Federal Office for Consumer Protection and Food Safety (Bundesamt für Verbraucherschutz und Lebensmittelsicherheit — BVL) according to the German Food and Feed Code, Section 64 (Lebensmittel- und Futtermittelgesetzbuch-LFGB, § 64) in accordance with ISO 5725-2, -4 and -6 with 16 participating laboratories from different European countries (including 7 NRLs).

This validation study was carried out in three independent sub-studies. Precision data for okadaic acid, DTX-2, DTX-1 (including their esters), AZA-1, AZA-2, AZA-3, YTX, 45-OH-YTX and PTX-2 were determined in naturally contaminated blue mussels (homogenate of cooked and raw mussels), spiked extracts of oysters (homogenate of raw shellfish) and spiked extracts of clams (homogenate of raw shellfish). Test samples 1 to 8 were analyzed by 13 laboratories by means of duplicate injection (analogous to the repeatability standard deviation (SD) of an extract) in three separate sequences (intermediate conditions). Test samples 9 to 12 were analyzed by eight laboratories within one single sequence by means of duplicate injection (analogous to the repeatability standard deviation (SD) of an extract). Test samples 13 to 16 were analysed by 15 laboratories within one single sequence by means of duplicate injection (analogous to the repeatability standard deviation (SD) of an extract). The results of the statistical evaluation of the investigated toxins are given in Table A.1 to Table A.12. The corrected HORRAT values were calculated according to [17].

The different analytical conditions, i.e. acidic and basic mobile phases, have been used by the different participants and had no influence on the laboratory results.

Sample 2 is a blank blue mussel homogenate which was found by all laboratories as an uncontaminated material.

Table A.1 — Precision data for Azaspiracid-1 (AZA-1)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13		13	13	13	7	7	15	15		
Number of laboratories retained after eliminating outliers	13		11	12	13	7	7	15	15		
Number of outliers (laboratories)	0		2	1	0	0	0	0	0		
Number of accepted results	156		66	72	156	28	28	30	30		
Mean value, \bar{x} , $\mu\text{g/kg}$	96,36	< LOD	111,7	82,5	99,5	17,4	71,9	127,54	64,92	< LOD	< LOD
Repeatability standard deviation s_r , $\mu\text{g/kg}$	11,4		8,9	8,2	10,4	1,8	3,5				
Repeatability relative standard deviation, RSD_r , %	11,8		8,0	9,9	10,5	10,4	4,9				
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$	31,8		25,0	22,9	29,2	5,1	9,9				
Reproducibility standard deviation s_R , $\mu\text{g/kg}$	20,4		25,9	12,1	22,5	6,6	22,1	19,4	10,6		
Reproducibility relative standard deviation, RSD_R , %	21,2		23,2	14,7	22,6	37,9	30,8	15,2	16,3		
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$	57,2		72,6	34,0	63,0	18,5	62,0	54,3	29,6		
Intermediate conditions, %	14,1		9,6	10,7	11,9						
HorRat value, corrected according to [17]	0,96		1,05	0,67	1,03	1,72	1,40	0,69	0,74		
HorRat _R value, uncorrected	0,93		1,04	0,63	1,00	1,29	1,30	0,69	0,67		

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

Table A.2 — Precision data for Azaspiracid-2 (AZA-2)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13		13				7	15	14		
Number of laboratories retained after eliminating outliers	13		13				6	14	13		
Number of outliers (laboratories)	0		0				1	1	1		
Number of accepted results	156		78				24	28	26		
Mean value, \bar{x} , $\mu\text{g/kg}$	21,80	< LOD	28,90	< LOD	< LOD	< LOD	10,78	36,22	19,29	< LOD	< LOD
Repeatability standard deviation s_r , $\mu\text{g/kg}$	3,16		4,7				1,0				
Repeatability relative standard deviation, RSD_r , %	14,5		16,3				9,5				
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$	8,85		13,2				2,9				
Reproducibility standard deviation s_R , $\mu\text{g/kg}$	8,24		7,6				4,5	11,9	6,6		
Reproducibility relative standard deviation, RSD_R , %	37,8		26,2				42,0	32,8	34,0		
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$	23,07		21,2				12,7	33,3	18,4		
Intermediate conditions, %	24,4		18,0								
HorRat value, corrected according to [17]	1,72		1,19				1,91	1,49	1,54		
HorRat _R value, uncorrected	1,33		0,96				1,33	1,24	1,17		
^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate. ^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2. ^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX. ^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate). ^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.											

Table A.3 — Precision data for Azaspiracid-3 (AZA-3)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13		13			7	7	13	9		
Number of laboratories retained after eliminating outliers	13		12			6	6	12	8		
Number of outliers (laboratories)	0		1			1	1	1	1		
Number of accepted results	156		72			24	24	24	16		
Mean value, \bar{x} , $\mu\text{g/kg}$	50,27	< LOD	49,88	< LOD	< LOD	10,98	41,59	11,02	8,11	< LOD	< LOD
Repeatability standard deviation s_r , $\mu\text{g/kg}$	5,5		6,3			0,7	1,9				
Repeatability relative standard deviation, RSD_r , %	10,9		12,7			6,1	4,6				
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$	15,3		17,7			1,9	5,4				
Reproducibility standard deviation s_R , $\mu\text{g/kg}$	14,7		12,3			3,6	11,4	3,7	1,8		
Reproducibility relative standard deviation, RSD_R , %	29,2		24,6			32,4	27,3	33,7	22,4		
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$	41,1		34,4			10,0	31,8	10,4	5,1		
Intermediate conditions, %	13,1		13,7								
HorRat value, corrected according to [17]	1,33		1,12			1,47	1,24	1,53	1,02		
HorRat _R value, uncorrected	1,16		0,98			1,03	1,06	1,07	0,68		

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

Table A.4 — Precision data for Okadaic acid (OA)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13	13	13	13	13	8	8	15	15	15	15
Number of laboratories retained after eliminating outliers	13	11	12	12	13	7	8	15	14	15	14
Number of outliers (laboratories)	0	2	1	1	0	1	0	0	1	0	1
Number of accepted results	156	66	72	72	156	28	32	30	28	30	28
Mean value, \bar{x} , $\mu\text{g/kg}$	23,32	36,71	23,07	130,79	81,79	17,94	18,85	111,04	52,45	37,55	25,66
Repeatability standard deviation s_r , $\mu\text{g/kg}$	5,7		5,1	18,7	10,7	2,3	2,6				
Repeatability relative standard deviation, RSD_r , %	24,6		22,2	14,3	13,1	12,9	13,9				
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$	16,1		14,3	52,4	30,0	6,5	7,3				
Reproducibility standard deviation s_R , $\mu\text{g/kg}$	6,9	9,0	6,9	19,1	17,0	6,1	6,2	32,0	14,8	10,1	5,8
Reproducibility relative standard deviation, RSD_R , %	29,4	24,6	30,1	14,6	20,8	33,9	32,6	28,8	28,3	26,9	22,7
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$	19,2	25,3	19,4	53,5	47,6	17,0	17,2	89,6	41,6	28,3	16,3
Intermediate conditions, %	29,0	24,6	20,8	14,3	17,3						
HorRat value, corrected according to [17]	1,34	1,12	1,37	0,67	0,95	1,54	1,48	1,31	1,28	1,22	1,03
HorRat _R value, uncorrected	1,04	0,94	1,07	0,67	0,89	1,16	1,12	1,29	1,13	1,02	0,82
^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate. ^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2. ^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX. ^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate). ^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.											

Table A.5 — Precision data for Okadaic acid (OA) after hydrolysis

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13	13						15	15	15	15
Number of laboratories retained after eliminating outliers	13	13						13	13	13	13
Number of outliers (laboratories)	0	0						2	2	2	2
Number of accepted results	156	78						26	26	26	26
Mean value, \bar{x} , µg/kg	48,08	78,57	< LOD	< LOD	< LOD	< LOD	< LOD	174,67	79,37	65,53	44,09
Repeatability standard deviation s_r , µg/kg	9,5										
Repeatability relative standard deviation, RSD_r , %	19,7										
Repeatability limit r [$r = 2,8 \times s_r$], µg/kg	26,5										
Reproducibility standard deviation s_R , µg/kg	14,0	20,4						54,0	21,3	17,6	10,9
Reproducibility relative standard deviation, RSD_R , %	29,2	25,9						30,9	26,8	26,9	24,8
Reproducibility limit R [$R = 2,8 \times s_R$], µg/kg	39,3	57,0						151,1	59,6	49,4	30,6
Intermediate conditions, %	25,7	19,4									
HorRat value, corrected according to [17]	1,33	1,18						1,48	1,22	1,22	1,13
HorRat _R value, uncorrected	1,16	1,10						1,48	1,14	1,11	0,97

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

Table A.6 — Precision data for Dinophysistoxin-1 (DTX-1)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories		13				8	8			15	15
Number of laboratories retained after eliminating outliers		13				7	7			14	14
Number of outliers (laboratories)	0	0				1	1			1	1
Number of accepted results		78				28	28			28	28
Mean value, \bar{x} , $\mu\text{g/kg}$	< LOD	68,51	< LOD	< LOD	< LOD	80,26	33,17	< LOD	< LOD	219,17	118,11
Repeatability standard deviation s_r , $\mu\text{g/kg}$						5,1	5,7				
Repeatability relative standard deviation, RSD_r , %						6,3	17,3				
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$						14,2	16,1				
Reproducibility standard deviation s_R , $\mu\text{g/kg}$		23,3				27,1	12,8			76,7	42,3
Reproducibility relative standard deviation, RSD_R , %		34,0				33,8	38,7			35,0	35,8
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$		65,2				76,0	35,9			214,8	118,4
Intermediate conditions, %		20,6									
HorRat value, corrected according to [17]		1,55				1,54	1,76			1,74	1,63
HorRat _R value, uncorrected		1,42				1,45	1,45			1,74	1,62

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

Table A.7 — Precision data for Dinophysistoxin-1 (DTX-1) after hydrolysis

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories		13								15	15
Number of laboratories retained after eliminating outliers		13								14	14
Number of outliers (laboratories)		0								1	1
Number of accepted results		78								28	28
Mean value, \bar{x} , µg/kg	< LOD	96,54	< LOD	< LOD	< LOD	< LOD	< LOD	< LOD	< LOD	424,09	221,44
Repeatability standard deviation s_r , µg/kg											
Repeatability relative standard deviation, RSD_r , %											
Repeatability limit r [$r = 2,8 \times s_r$], µg/kg											
Reproducibility standard deviation s_R , µg/kg		31,5								149,7	72,6
Reproducibility relative standard deviation, RSD_R , %		32,6								35,3	32,8
Reproducibility limit R [$R = 2,8 \times s_R$], µg/kg		88,12								419,2	203,4
Intermediate conditions, %		22,0									
HorRat value, corrected according to [17]		1,48								1,94	1,63
HorRat _R value, uncorrected		1,43								1,94	1,63

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

Table A.8 — Precision data for Dinophysistoxin-2 (DTX-2)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13	13	13			8	8	15	15		
Number of laboratories retained after eliminating outliers	13	13	11			8	8	15	15		
Number of outliers (laboratories)	0	0	2			0	0	0	0		
Number of accepted results	156	78	66			32	32	30	30		
Mean value, \bar{x} , $\mu\text{g/kg}$	129,50	119,74	116,52	< LOD	< LOD	36,21	82,02	187,95	99,20	< LOD	< LOD
Repeatability standard deviation s_r , $\mu\text{g/kg}$	19,29		14,10			6,95	16,65				
Repeatability relative standard deviation, RSD_r , %	14,9		12,1			19,2	20,3				
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$	54,03		39,48			19,47	46,62				
Reproducibility standard deviation s_R , $\mu\text{g/kg}$	49,73	40,59	40,90			9,78	31,99	40,4	27,3		
Reproducibility relative standard deviation, RSD_R , %	38,4	33,9	35,1			27,0	39,0	21,5	27,5		
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$	139,24	113,66	114,51			27,37	89,56	113,1	76,4		
Intermediate conditions, %	19,8	18,5	13,1								
HorRat value, corrected according to [17]	1,76	1,54	1,60			1,23	1,77	1,05	1,25		
HorRat _R value, uncorrected	1,76	1,54	1,59			1,02	1,67	1,05	1,21		
^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate. ^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2. ^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX. ^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate). ^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates;											

Table A.9 — Precision data for Dinophysistoxin-2 (DTX-2) after hydrolysis

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories	13	13									
Number of laboratories retained after eliminating outliers	13	13									
Number of outliers (laboratories)	0	0									
Number of accepted results	156	78									
Mean value, \bar{x} , $\mu\text{g/kg}$	179,42	155,43	< LOD	< LOD	< LOD	< LOD	< LOD	203,78	104,14	< LOD	< LOD
Repeatability standard deviation s_r , $\mu\text{g/kg}$	26,0										
Repeatability relative standard deviation, RSD_r , %	14,5										
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$	72,8										
Reproducibility standard deviation s_R , $\mu\text{g/kg}$	51,0	43,5						68,3	37,2		
Reproducibility relative standard deviation, RSD_R , %	28,4	28,0						33,5	35,7		
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$	142,7	121,9						191,1	104,1		
Intermediate conditions, %	16,8	13,8									
HorRat value, corrected according to [17]	1,37	1,32						1,65	1,62		
HorRat _R value, uncorrected	1,37	1,32						1,65	1,59		

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates;

Table A.10 — Precision data for Yessotoxin (YTX)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories		12		12	12					14	14
Number of laboratories retained after eliminating outliers		11		12	10					13	13
Number of outliers (laboratories)		1		0	2					1	1
Number of accepted results		66		72	120					26	26
Mean value, \bar{x} , $\mu\text{g/kg}$	< LOD	311,80	< LOD	367,64	213,69	< LOD	< LOD	< LOD	< LOD	471,40	624,41
Repeatability standard deviation s_r , $\mu\text{g/kg}$				43,4	16,0						
Repeatability relative standard deviation, RSD_r , %				11,8	8,4						
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$				121,5	50,3						
Reproducibility standard deviation s_R , $\mu\text{g/kg}$		45,8		48,5	31,6					104,6	151,0
Reproducibility relative standard deviation, RSD_R , %		14,7		13,2	14,8					22,2	24,2
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$		128,3		135,9	88,6					293,0	422,9
Intermediate conditions, %		14,7		11,8	11,3						
HorRat value, corrected according to [17]		0,77		0,71	0,73					1,24	1,41
HorRat _R value, uncorrected		0,77		0,71	0,73					1,24	1,41
^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate. ^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2. ^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX. ^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate). ^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate). ^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate); ^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.											

Table A.11 — Precision data for 45-Hydroxy-Yessotoxin (45-OH-YTX)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories		6								10	10
Number of laboratories retained after eliminating outliers		6								9	9
Number of outliers (laboratories)		0								1	1
Number of accepted results		36								18	18
Mean value, \bar{x} , $\mu\text{g/kg}$	< LOD	212,22	< LOD	< LOD	< LOD	< LOD	< LOD	< LOD	< LOD	345,76	463,75
Repeatability standard deviation s_r , $\mu\text{g/kg}$											
Repeatability relative standard deviation, RSD_r , %											
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$											
Reproducibility standard deviation s_R , $\mu\text{g/kg}$		88,7								103,4	155,8
Reproducibility relative standard deviation, RSD_R , %		41,8								29,9	33,6
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$		248,4								289,5	436,3
Intermediate conditions, %		20,2									
HorRat value, corrected according to [17]		2,07								1,59	1,87
HorRat _R value, uncorrected		2,07								1,59	1,87

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

Table A.12 — Precision data for Pectenotoxin-2 (PTX-2)

Sample	1/3 ^a	4 ^b	5 ^c	7 ^d	6/8 ^e	9/11 ^f	10/12 ^g	13 ^h	14 ⁱ	15 ^j	16 ^k
Year of inter-laboratory test	2009	2009	2009	2009	2009	2009	2009	2010	2010	2010	2010
Number of laboratories			12	12	12					15	15
Number of laboratories retained after eliminating outliers			12	11	12					15	15
Number of outliers (laboratories)			0	1	0					0	0
Number of accepted results			72	66	72					30	30
Mean value, \bar{x} , $\mu\text{g/kg}$	< LOD	< LOD	217,57	203,40	117,89	< LOD	< LOD	< LOD	< LOD	26,75	14,33
Repeatability standard deviation s_r , $\mu\text{g/kg}$			15,7	16,3	12,3						
Repeatability relative standard deviation, RSD_r , %			7,2	8,0	10,4						
Repeatability limit r [$r = 2,8 \times s_r$], $\mu\text{g/kg}$			43,9	45,6	34,3						
Reproducibility standard deviation s_R , $\mu\text{g/kg}$			33,5	44,5	16,0					8,9	5,0
Reproducibility relative standard deviation, RSD_R , %			15,4	21,9	13,6					33,3	35,0
Reproducibility limit R [$R = 2,8 \times s_R$], $\mu\text{g/kg}$			93,8	124,7	44,9					24,9	14,0
Intermediate conditions, %			10,6	10,0	10,5						
HorRat value, corrected according to [17]			0,77	1,08	0,62					1,52	1,59
HorRat _R value, uncorrected			0,77	1,08	0,62					1,21	1,16

^a Samples 1/3: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^b Sample 4: Blue mussels, cooked, naturally contaminated homogenate.

^c Sample 5: Extract of samples 1/3 and additionally spiked with PTX-2.

^d Sample 7: Clams, raw, extract of blank clams spiked with OA, AZA-1, PTX-2 and YTX.

^e Samples 6/8: Oysters, raw, extract of blank oysters spiked with OA, AZA-1, PTX-2 and YTX (blind duplicate).

^f Samples 9/11: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^g Samples 10/12: Blue mussels, cooked, naturally contaminated homogenate (blind duplicate).

^{h to k} Samples 13 to 16: Blue mussels, raw, naturally contaminated homogenates.

A.2 Recovery

The overall recovery in LC-MS/MS procedures is the sum of the extraction recovery and matrix effects induced during analysis. The extraction recovery of OA, PTX-2, AZA-1 and YTX were determined for blue mussels (cooked shellfish), oysters (raw shellfish) and clams (raw shellfish) at concentration level of 50 µg/kg for OA, PTX-2, AZA-1 and at concentration level of 100 µg/kg for YTX. The analyses for the determination of the extraction recovery were carried out within the organizing laboratory. It was shown during the collaborative study that the extraction efficiency for blue mussels of the organizing laboratory is representative of the other participating laboratories. The values are summarized in Table A.13 and are given as mean with standard deviations (see [8]).

Table A.13

	OA	YTX	AZA-1	PTX-2
Extraction recovery in % (Blue mussels, cooked), <i>n</i> = 4	98 ± 13	81 ± 8	85 ± 5	99 ± 17
Extraction recovery in % (Clam, raw), <i>n</i> = 4	73 ± 2	90 ± 14	72 ± 4	81 ± 4
Extraction recovery in % (Oysters, raw), <i>n</i> = 4	78 ± 1	85 ± 6	91 ± 4	90 ± 1

The laboratory-comprehensive recovery values obtained after matrix correction from spiked extracts (Sample 5, 6/8 and 7) were collected during the collaborative study and are summarized in Table A.14. The values are given as mean with standard deviations. Apart from two cases concerning PTX-2 in Sample 5 and YTX in Sample 7, no significant deviations from 100% were notable, reflecting a sufficient accuracy of LC-MS/MS measurement. This also supports the concept of determination of apparent recoveries for a single shellfish species and the application of this matrix effect correction to other shellfish species (see [8]).

Table A.14

	OA	YTX	AZA-1	PTX-2
Laboratory-comprehensive recovery rate in % (Sample 5, Blue mussels, cooked)	< LOD	< LOD	< LOD	86±8 (<i>n</i> =12)
Laboratory-comprehensive recovery rate in % (Sample 7, Clams, raw)	105 ± 9 (<i>n</i> =12)	91 ± 6 (<i>n</i> =12)	100 ± 11 (<i>n</i> =12)	92 ± 9 (<i>n</i> =11)
Laboratory-comprehensive recovery rate in % (Sample 6/8, Oysters, raw)	102 ± 10 (<i>n</i> =13)	95 ± 9 (<i>n</i> =10)	99 ± 10 (<i>n</i> =13)	97 ± 8 (<i>n</i> =12)

Annex B (informative)

Examples for suitable MS detection conditions

B.1 Examples suitable for SCIEX API 4000 or API 4000 Q-Trap

Source parameters and general mass spectrometric specifications (see also Table B.1):

- Ionization: Electrospray ionization (ESI)
- Temperature: 400 °C to 500 °C
- Curtain gas: 14,0 psi
- Collision gas: 9 psi/medium
- Entrance potential (EP): 10 V
- Ion spray voltage: for positive mode: 5 000 V
for negative mode: 4 500 V
- Ion source gas 1 and 2: 35 psi (negative mode)
- Ion source gas 1 and 2: 50 psi (positive mode)
- Q1 Resolution: unit
- Q3 Resolution: unit

Table B.1 — Substance-specific mass spectrometric specifications

Substance	Polarity	Q1 Mass	Q3 Mass	Dwell [msec]	DP ^a [V]	CE ^b [eV]	CXP ^c [V]
OA and DTX-2	neg	803,4	255,1	100	−145	−62	16
	neg	803,4	113,1	100	−145	−92	10
	pos	805,5	769,5	100	76	21	26
OA/DTX-2 (Na Adduct)	pos	826,5	723,5	100	120	65	26
DTX-1	neg	816,3	255,4	100	−145	−62	16
	neg	816,3	113,2	100	−180	−92	10
	pos	819,5	783,5	100	76	21	26
DTX-1 (Na Adduct)	pos	841,5	736,8	100	120	65	26
PTX-2 (NH ₄ Adduct)	pos	876,5	823,5	100	150	+27	15
PTX-2 (NH ₄ Adduct)	pos	876,5	805,5	100	150	+50	15
	neg	856,41	626,4	100	−85	−50	29
	neg	856,41	136,9	100	−150	−80	9
PTX-1 (NH ₄ Adduct)	pos	892,5	874,5	100			
PTX-1 (NH ₄ Adduct)	pos	892,5	856,5	100			
PTX-2 sa	pos	894,5	859,5	100	150	+27	15
YTX	neg	1 141,6	1 061,7	100	−150	−42	25
45-OH-YTX	neg	1 156,5	1 076,5	100	−150	−42	26
Homo-YTX	neg	1 155,5	1 075,5	100	−150	−42	27
45-OH-Homo-YTX	neg	1 171,5	1 091,5	100	−150	−42	25
AZA-1	pos	842,62	824,60	100	100	+40	44
	pos	842,62	806,3	100	100	+55	24
	pos	842,62	672,4	100	100	+70	44
	pos	842,62	654,4	100	100	+70	24
AZA-2	pos	856,5	838,5	100	100	+40	44
	pos	856,5	820,4	100	100	+55	24
	pos	856,5	672,4	100	100	+70	44
	pos	856,5	654,4	100	100	+70	24
AZA-3	pos	828,5	810,5	100	100	+40	44
	pos	828,5	792,4	100	100	+55	24
	pos	828,5	658,4	100	100	+70	44
	pos	828,5	640,4	100	100	+70	24
^a Declustering potential. ^b Collision energy. ^c Cell exit potential.							

B.2 Examples suitable for Waters (Micromass) TSQ Ultima

Source parameters and general mass spectrometric specifications (see also Table B.2):

— Ionization:	Electrospray ionization (ESI)
— Source temperature:	150 °C
— Desolvation temperature:	350 °C
— Cone gas flow:	23 l/h
— Cone:	50 V
— Collision gas:	1,95 mbar
— Hexapole:	30 V to 45 V
— Ion spray voltage:	for positive mode: 3 300 V
	for negative mode: 3 500 V
— Ion energy1:	1,2
— Ion energy1:	2,0
— Low mass (LM) resolution:	13,0
— High mass (HM) resolution:	13,0

Table B.2 — Substance-specific mass spectrometric specifications

Substance	Polarity	Q1 Mass	Q3 Mass	Dwell [msec]	Cone voltage [V]	CE ^a [eV]
OA and DTX-2	neg	803,5	255,1	100	−77	−40
	neg	803,5	803,5	100	−77	−40
DTX-1	neg	817,5	255,1	100	−77	−40
	neg	817,5	817,5	100	−77	−40
PTX-2 (NH ₄ Adduct)	pos	876,5	823,5	100	30	30
	pos	876,5	805,5	100	30	30
YTX	neg	1 141,6	1 061,6	100	−35	−30
	neg	1 141,6	1 141,6	100	−35	−30
45-OH-YTX	neg	1 157,5	1 077,5	100	−35	−30
Homo-YTX	neg	1 155,5	1 075,5	100	−35	−30
45-OH-Homo-YTX	neg	1 171,5	1 091,5	100	−35	−30
AZA-1	pos	842,5	654,4	100	60	+42
	pos	842,5	672,4	100	60	+42
AZA-2	pos	856,5	654,4	100	60	+42
	pos	856,5	672,4	100	60	+42
AZA-3	pos	828,5	810,5	100	60	+42
	pos	828,5	640,4	100	60	+42
^a Collision energy.						

B.3 Examples suitable for Thermo Fisher TSQ Quantum Ultra

Source parameters and general mass spectrometric specifications (see also Table B.3):

— Ionization:	Electrospray ionization (ESI)
— Temperature:	250 °C
— Sheath gas:	30
— Collision gas:	1,5 mtorr
— Ion spray voltage:	for positive mode: 5 000 V for negative mode: no data available
— Auxillary gas:	0
— Q1 Resolution:	0,7
— Q3 Resolution:	0,7

Table B.3 — Substance-specific mass spectrometric conditions

Substance	Polarity	Q1 Mass	Q3 Mass	Dwell [msec]	Skimmer offset [V]	CE ^a [eV]
OA and DTX-2_(Na Adduct)	pos	827,45	723,35	100	10	48
	pos	827,45	809,4	100	10	43
OA and DTX-2_(NH ₄ -Adduct)	pos	822,45	751,4	100	10	20
	pos	822,45	769,4	100	10	15
DTX-1_(Na Adduct)	pos	841,45	737,35	100	10	48
	pos	841,45	823,5	100	10	43
DTX-1_(NH ₄ Adduct)	pos	836,45	765,4	100	10	20
	pos	836,45	783,4	100	10	15
PTX-2_(NH ₄ Adduct)	pos	876,45	823,45	100	10	27
	pos	876,45	805,45	100	10	30
PTX-2_(Na Adduct)	pos	857,41	837,45	100	10	55
	pos	857,41	539,3	100	10	60
PTX-1 (NH ₄ Adduct)	pos	892,5	874,5	100		
PTX-1 (NH ₄ Adduct)	pos	892,5	856,5	100		
PTX-2 sa	pos	894,5	859,5	100		
YTX	neg	1 141,6	1 061,7	100		
45-OH-YTX	neg	1 157,5	1 077,5	100		
Homo-YTX	neg	1 155,5	1 075,5	100		
45-OH-Homo-YTX	neg	1 171,5	1 091,5	100		
AZA-1	pos	842,45	824,45	100	10	30
	pos	842,45	672,4	100	10	40
AZA-2	pos	856,5	838,45	100	10	30
	pos	856,5	672,4	100	10	40
AZA-3	pos	828,5	810,45	100	10	30
	pos	828,5	658,4	100	10	40
^a Collision energy.						

B.4 Examples suitable for Agilent 6410 or 6460 QQQ

Source parameters and general mass spectrometric conditions (see also Table B.4):

— Ionization:	Electrospray ionization (ESI)
— Gas temp:	300 °C
— Gas flow:	10 l/min
— Nebulizer:	43 psi
— Capillary:	for positive mode: 4 500 V
	for negative mode: 5 200 V
— Delta EMV:	400

ESI + Agilent JetStream:


— Gas temp:	300 °C
— Gas flow:	5 l/min
— Nebulizer:	45 psi
— Sheath gas temp:	250 °C
— Sheath gas flow:	11 l/min
— Capillary:	for positive mode: 3 500 V
	for negative mode: 3 500 V
— Nozzle voltage:	500 V
— Delta EMV:	400

Table B.4— Substance-specific mass spectrometric conditions

Substance	Polarity	Prec Ion	Prod Ion	Dwell [msec]	Frag [V]	CE ^a [V]
OA and DTX-2 (Na Adduct)	pos	827,5	809,2	25	220	45
	pos	827,5	723,4	25	220	55
DTX-1 (Na Adduct)	pos	841,5	823,2	25	220	45
	pos	841,5	737,2	25	220	55
PTX-2 (Na Adduct)	pos	881,5	837,5	25	230	60
	pos	881,5	539,3	25	230	70
PTX-1 (Na Adduct)	pos	897,5	853,5	25	230	60
	pos	897,5	555,3	25	230	70
<i>PTX-2 sa</i> (Na Adduct)	pos	899,5	855,5	25	230	60
	pos	899,5	557,3	25	230	70
YTX	neg	1 141,5	1 061,3	100	135	35
	neg	1 141,5	925,5	100	135	60
45-OH-YTX	neg	1 157,4	1 077,5	100	135	35
Homo-YTX	neg	1 155,4	1 075,5	100	135	35
45-OH-Homo-YTX	neg	1 171,4	1 091,5	100	135	35
AZA-1	pos	842,5	824,5	25	200	40
	pos	842,5	806,5	25	200	55
AZA-2	pos	856,5	838,5	25	200	40
	pos	856,5	820,5	25	200	55
AZA-3	pos	828,5	810,5	25	200	40
	pos	828,5	792,5	25	200	55
^a Collision energy.						

Annex C (informative)

Typical chromatogram



Key

Y1: relative intensity;

Y2: portion of eluent B₂ in %;

X: time, in min;

Black line: negative mode

Grey line: positive mode

(SCIEX API 4000 Q-trap; basic chromatographic (7.3), using Phenomenex Gemini NX C18 (150 mm x 2 mm, 3 µm), relative intensity of PTX-2 is increased to the factor 20)

Figure C.1 — Typical chromatogram of a laboratory reference material

Bibliography

- [1] L.A. Stobo, J.P.C.L. Lacaze, A.C. Scott, S. Gallacher, E.A. Smith, M.A. Quilliam, 2005. *Liquid chromatography with mass spectrometry — Detection of lipophilic shellfish toxins*, J. AOAC Int. 88, 1371-1382
- [2] P. McNabb, A.I. Selwood, P.T. Holland, 2005. *Multiresidue method for determination of algal toxins in shellfish: Single-laboratory validation and interlaboratory study*, J. AOAC Int. 88, 761-772
- [3] A. These, J. Scholz, A. Preiss-Weigert, 2009. *Sensitive method for the determination of lipophilic marine biotoxins in extracts of mussels and processed shellfish by high-performance liquid chromatography-tandem mass spectrometry based on enrichment by solid-phase extraction*, J. of Chromatography A, 1216, 4529-4538
- [4] *Quantitative determination of marine lipophilic toxins in shellfish using LC-MS/MS*, Final Report, February 2011, RIKILT – Institute of Food Safety, Wageningen (NL)
- [5] *Inter-laboratory Validation Study of the “EU-Harmonised Standard Operation Procedure for determination of Lipophilic marine biotoxins in molluscs by LC-MS/MS”*, Draft Report, European Reference Laboratory for Marine Biotoxins, Vigo (ES)
- [6] D.O. Mountfort, T. Suzuki, P. Truman, 2001. *Protein phosphatase inhibition assay adapted for determination of total DSP in contaminated mussels*, Toxicon 39, 383-390
- [7] Hess, P., McCarron, P., Quilliam, M. A. 2007. *Fit-for-purpose shellfish reference materials for internal and external quality control in the analysis of phycotoxins*, Anal Bioanal Chem. 387 (8), 2463-2474
- [8] These A, Klemm C, Nausch I, Uhlig S. 2011. *Results of a European interlaboratory method validation study for the quantitative determination of lipophilic marine biotoxins in raw and cooked shellfish based on high-performance liquid chromatography-tandem mass spectrometry. Part I: collaborative study*, Anal Bioanal Chem. 399 (3); 1245-1256
- [9] *Marine biotoxins in shellfish – Okadaic acid and analogues; Scientific Opinion of the Panel on Contaminants in the Food chain*; The EFSA Journal (2008) 589, 1-62 in its current version
- [10] *Marine biotoxins in shellfish – Azaspiracid group; Scientific Opinion of the Panel on Contaminants in the Food chain*; The EFSA Journal (2008) 723, 1-52 in its current version
- [11] *Marine biotoxins in shellfish – Pectenotoxin group; Scientific Opinion of the Panel on Contaminants in the Food chain*; The EFSA Journal (2009) 1109, 1-47 in its current version
- [12] *Marine biotoxins in shellfish – Yessotoxin group; Scientific Opinion of the Panel on Contaminants in the Food chain*; The EFSA Journal (2008) 907, 1-62 in its current version
- [13] ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [14] ISO 5725-4:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 4: Basic methods for the determination of the trueness of a standard measurement method*
- [15] ISO 5725-6:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 6: Use in practice of accuracy values*
- [16] Amtl. Sammlung von Untersuchungsverfahren nach § 64 LFGB (Official Collection of Testing Methods according to § 64 of the Foodstuffs and Animal Feed Code – LFGB), Band 1 (Volume 1), L 12.03/04-4,

Untersuchung von Lebensmitteln, Bestimmung von lipophilen Algentoxinen (DSP-Toxine, Yessotoxine, Azaspirosäuren, Pectenotoxine) in Muscheltieren und Muscheltiererzeugnissen LC-MS/MS-Verfahren (Determination of Lipophilic Algal Toxins (DSP-Toxins, Yessotoxins, Azaspiracids, Pectenotoxins) in shellfish and shellfish products by LC-MS/MS) September 2010

- [17] Thompson, M. and Wood, R., 2000, *Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing*, Analyst 125, 385-386
- [18] Commission Regulation 15/2011 of 10 January 2011 amending Regulation (EC) No 2074/2005 as regards recognised testing methods for detecting marine biotoxins in live bivalve molluscs

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com