



BSI Standards Publication

Animal and vegetable fats and oils — Determination of melting point in open capillary tubes — Slip point

National foreword

This British Standard is the UK implementation of EN ISO 6321:2021. It is identical to ISO 6321:2021. It supersedes BS EN ISO 6321:2002, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee AW/307, Oilseeds, animal and vegetable fats and oils and their by-products.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2021
Published by BSI Standards Limited 2021

ISBN 978 0 539 14633 2

ICS 67.200.10; 83.120

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2021.

Amendments/corrigenda issued since publication

| Date | Text affected |
|------|---------------|
|------|---------------|

EUROPEAN STANDARD

EN ISO 6321

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2021

ICS 67.200.10

Supersedes EN ISO 6321:2002

English Version

**Animal and vegetable fats and oils - Determination
of melting point in open capillary tubes - Slip point
(ISO 6321:2021)**

Corps gras d'origines animale et végétale
- Détermination du point de fusion en tube
capillaire ouvert (ISO 6321:2021)

Tierische und pflanzliche Fette und Öle -
Bestimmung des Schmelzpunktes in offenen
Kapillarröhrchen (ISO 6321:2021)

This European Standard was approved by CEN on 18 May 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 6321:2021) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 307 "Oilseeds, vegetable and animal fats and oils and their by-products - Methods of sampling and analysis" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2021, and conflicting national standards shall be withdrawn at the latest by December 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 6321:2002.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 6321:2021 has been approved by CEN as EN ISO 6321:2021 without any modification.

Contents

Page

| | |
|---|-----------|
| Foreword | iv |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 Principle | 1 |
| 5 Apparatus | 2 |
| 6 Sampling | 4 |
| 7 Preparation of test sample | 4 |
| 8 Procedure | 5 |
| 8.1 Preparation of the capillary tubes for method A | 5 |
| 8.2 Preparation of the capillary tubes for method B | 5 |
| 8.3 Determination | 5 |
| 8.4 Number of determinations | 6 |
| 9 Expression of results | 6 |
| 10 Precision | 6 |
| 10.1 Interlaboratory tests | 6 |
| 10.2 Repeatability | 6 |
| 11 Test report | 6 |
| Annex A (normative) Method for palm oil samples | 7 |
| Annex B (informative) Results of interlaboratory tests | 8 |
| Bibliography | 11 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 307, *Oilseeds, vegetable and animal fats and oils and their by-products — Methods of sampling and analysis*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition ([ISO 6321:2002](http://www.iso.org/iso/6321:2002)), which has been technically revised.

The main changes compared to the previous edition are as follows:

- the requirement to measure the diameters of each capillary tube has been removed, and
- a footnote stating suggested suppliers of suitable capillary tubes has been included.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Animal and vegetable fats and oils — Determination of melting point in open capillary tubes — Slip point

1 Scope

This document specifies two methods for the determination of the melting point in open capillary tubes, commonly known as the slip melting point, of animal and vegetable fats and oils (referred to as fats hereinafter).

- Method A is only applicable to animal and vegetable fats which are solid at ambient temperature and which do not exhibit pronounced polymorphism.
- Method B is applicable to all animal and vegetable fats which are solid at ambient temperature and is the method to be used for fats whose polymorphic behaviour is unknown.

For the determination of the slip melting point of palm oil samples the method given in [Annex A](#) shall be used.

NOTE 1 If applied to fats with pronounced polymorphism, method A will give different and less satisfactory results than method B.

NOTE 2 Fats which exhibit pronounced polymorphism are principally cocoa butter and fats containing appreciable quantities of 2-unsaturated, 1,3-saturated triacylglycerols.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

[ISO 661](#), *Animal and vegetable fats and oils — Preparation of test sample*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

slip melting point in open capillary tube

temperature at which a column of fat in an open capillary tube commences to rise under the conditions specified in this document

4 Principle

A capillary tube containing a column of the fat which has been crystallized under controlled conditions is immersed to a specified depth in water, the temperature of which is increased at a specified rate. The temperature at which the column is observed to start rising in the capillary tube is recorded.

5 Apparatus

Usual laboratory apparatus and, in particular, the following.

5.1 Capillary tubes, having uniform walls and which are open at both ends, of internal diameter 0,9 mm to 1,2 mm, external diameter 1,2 mm to 1,6 mm, wall thickness 0,15 mm to 0,30 mm and length 50 mm to 70 mm¹⁾.

Before use, clean the tubes thoroughly by washing them successively with a mixture of chromic acid, water and acetone or an alternative suitable cleaning solution, for example, hydrogen peroxide can be used. Dry the capillary tubes in an oven. However, it is recommended that new tubes are used.

5.2 Thermometer, graduated in divisions of 0,1 °C, calibrated over the range of melting points expected.

5.3 Stirrer, electrical.

5.4 Cooling bath, filled with brine or other non-freezing liquid, thermostatically maintained at a temperature of –10 °C to –12 °C, or filled with a mixture of flaked ice and salt (in the proportions 2 to 1 by mass) at a temperature of –10 °C to –12 °C.

5.5 Heating apparatus, consisting of the following elements:


- a) **water jacket**, made of glass, provided with inlet and outlet tubes, and having the shape and dimensions shown in [Figure 1](#);
- b) **water heater**, capable of delivering a slow stream of water, the temperature of which can be controlled to increase at a rate of between 0,5 °C/min and 4 °C/min, through the water jacket [[5.5 a](#)]).

An example of a suitable heating apparatus is shown in [Figure 2](#).

Other types of heating apparatus, such as a water bath with magnetic stirrer, capable of being controlled to produce the specified temperature rise [[5.5 b](#)]), may also be used.

1) Reference 1411022 from Hilgenberg (<https://www.hilgenberg-gmbh.de/innovative-glasprodukte/>), Reference 2930201 from Marienfeld (<https://www.marienfeld-superior.com/home.html>) and Reference 9201570 from Hirschmann (<http://www.hirschmann-laborgeraete.de/>) are examples of suitable products from providers of capillary tubes. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of products by these providers.


Dimensions in millimetres



Key
1 rubber band

Figure 1 — Water jacket

Dimensions in millimetres



Key

- 1 heating element (coil 220 W)
- a To drain.

Figure 2 — Example of heating apparatus (heating by natural convection)

6 Sampling

It is important the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

Sampling is not part of the method specified in this document. A recommended sampling method is given in [ISO 5555](#)^[1].

7 Preparation of test sample

Prepare the test sample in accordance with [ISO 661](#).

8 Procedure

8.1 Preparation of the capillary tubes for method A

Melt a portion of the test sample as rapidly as possible to at least 5 °C, but not more than 10 °C, above the temperature at which it is completely melted.

Dip two capillary tubes (5.1) into the melted test sample until columns of fat $10\text{ mm} \pm 2\text{ mm}$ long are obtained. Immediately after filling the tubes, wipe them quickly with absorbent tissue to remove any fat adhering to the outer surfaces of the tubes. Immediately place the filled capillary tubes for a few seconds against a beaker filled with ice so that the fat solidifies.

Place the tubes in the cooling bath (5.4) for 5 min.

Continue in accordance with 8.3.

8.2 Preparation of the capillary tubes for method B

Melt a portion of the test sample as rapidly as possible to at least 5 °C, but not more than 10 °C, above the temperature at which it is completely melted.

Cool the melted test sample, with occasional stirring, until its temperature is 32 °C to 34 °C and then stir continuously with the stirrer (5.3), allowing the fat to cool until the first signs of cloudiness appear.

Continue stirring by hand until the fat has a pasty consistency and then transfer the fat to a 100 ml beaker at $17\text{ °C} \pm 2\text{ °C}$.

Store the fat at this temperature for a minimum of 24 h.

Push four capillary tubes (5.1) into the conditioned fat until a column of fat $10\text{ mm} \pm 2\text{ mm}$ long is obtained in each tube. Wipe the tubes quickly with absorbent tissue to remove any fat adhering to the outer surfaces of the tubes.

Store the tubes at $17\text{ °C} \pm 2\text{ °C}$ until required.

8.3 Determination

8.3.1 Avoiding transfer of body heat to the fat, attach two capillary tubes prepared for method A (8.1) or for method B (8.2) to the thermometer (5.2) using small rubber bands (or by any other suitable means such as a rubber ring) so that the columns of fat are located at the lower ends of the tubes and lie adjacent to the bulb of the thermometer.

8.3.2 Fill the water jacket [5.5 a)] and the water heater [5.5 b)] with previously boiled water cooled to 15 °C. Clamp or suspend the thermometer with the attached capillary tubes centrally in the water jacket so that the lower ends of the capillary tubes are 30 mm below the surface of the water.

8.3.3 Operate the heating apparatus (5.5) so that a slow stream of water passes through the water jacket, regulating the heating so that the rise in temperature of the water, as measured by the thermometer in the water jacket, is about 3 °C/min to 4 °C/min for method A and 1 °C/min for method B.

8.3.4 For each of the two capillary tubes, note the temperature value indicated by the thermometer immediately as the fat starts to rise in the tube.

8.3.5 Note the arithmetic mean of the two readings obtained. For method A, take this arithmetic mean as the result of one determination.

8.3.6 For method B, repeat the operations described in [8.3.1](#) to [8.3.3](#) using the remaining two capillary tubes ([8.2](#)), decreasing the rate of temperature rise to 0,5 °C/min when the water temperature is within 5 °C of the mean reading determined in [8.3.5](#). For each of the two capillary tubes, note the temperature value indicated by the thermometer immediately the fat starts to rise in the tube. Record the arithmetic mean of the two readings obtained and take this as the result of one determination.

8.4 Number of determinations

Carry out two determinations on the same test sample [i.e. to obtain two mean readings for method A ([8.3.5](#)) and two final mean readings for method B ([8.3.6](#))].

9 Expression of results

Take as the result the arithmetic mean of the two determinations.

Express the slip melting point in open capillary tube to the nearest 0,1 °C.

10 Precision

10.1 Interlaboratory tests

Details of interlaboratory tests on the precision of the method are summarized in [Annex B](#). The values derived from these interlaboratory tests may not be applicable to concentration ranges and matrices other than those given.

10.2 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than 0,5 °C for method A and 1,0 °C for method B.

11 Test report

The test report shall specify:

- a) the sample;
- b) this document including year of publication (e.g. ISO 6321:2021);
- c) all information necessary for the complete identification of the sample;
- d) the sampling method used, if known;
- e) the test method used (i.e. ISO 6321, method A or method B);
- f) all operating details not specified in this document, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- g) the test result(s) obtained, or, if the repeatability has been checked, the final result obtained;
- h) any deviations from the procedure;
- i) any unusual features observed;
- j) the date of the test.

Annex A **(normative)**

Method for palm oil samples

Melt the sample and filter through a filter paper. Conduct the filtration in an oven set at 60 °C to avoid any crystallization of the sample. Leave the filtered sample in the oven for 10 min until it is free of air bubbles.

Dip at least three clean capillary tubes into the liquid sample so that columns of fat approximately 10 mm high are obtained in the tubes. Immediately chill the columns of fat by holding and rolling the ends of the tubes containing the sample pressed against a piece of ice, until the fat has solidified. Do not allow the open end of the tube to touch the ice. Wipe the tubes against a piece of tissue paper as quickly as possible. Place the tubes in a test tube which is held in a beaker of water that has been equilibrated at $10\text{ °C} \pm 1\text{ °C}$ in a thermostated water bath. Transfer the beaker to the water bath and hold for 16 h at $10\text{ °C} \pm 1\text{ °C}$.

For the determination, follow the procedure as given in [8.3.1](#) to [8.3.3](#). Regulate the rise in temperature in the water jacket to 1 °C/min, slowing down to 0,5 °C/min as the slip point is reached. Note the temperature value indicated by the thermometer as soon as the fat rises in each of the tubes.

Note the arithmetic mean of the three readings obtained and take this as the result of one determination.

Annex B (informative)

Results of interlaboratory tests

Two interlaboratory tests, carried out at the international level in 1982 and 1986 by ISO/TC 34/SC 11, in which 20 laboratories [each of which carried out three determinations on each sample (columns 2, 3 and 8)] and 15 laboratories [each of which carried out three determinations on each sample (columns 4 to 7)] participated, gave the statistical results (evaluated in accordance with [ISO 5725:1986^{2\)}](#)) shown in [Table B.1](#).

The results of interlaboratory tests on palm oil samples are given in [Tables B.2](#) and [B.3](#).

Table B.1 — Statistical results

| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 |
|--|-----------------|---------------------------|--------------|----------|--------------------------|-----------------------|-----------------------|
| | Method A | | Method B | | | | |
| | Palm kernel oil | Hydrogenated soyabean oil | Cocoa butter | Palm oil | Hydrogenated coconut oil | Hydrogenated palm oil | Hydrogenated palm oil |
| Number of laboratories retained after eliminating outliers | 18 | 18 | 14 | 14 | 13 | 13 | 18 |
| Mean (°C) | 27,6 | 35,4 | 31,4 | 36,3 | 37,1 | 45,5 | 47,5 |
| Standard deviation of repeatability, s_r (°C) | 0,15 | 0,14 | 0,29 | 0,35 | 0,30 | 0,13 | 0,15 |
| Coefficient of variation of repeatability (%) | 0,5 | 0,4 | 0,9 | 1,0 | 0,8 | 0,3 | 0,3 |
| Repeatability limit, r (2,8 s_r) (°C) | 0,4 | 0,4 | 0,8 | 1,0 | 0,8 | 0,4 | 0,4 |
| Standard deviation of reproducibility, s_R (°C) | 0,31 | 0,75 | 2,0 | 2,5 | 0,9 | 0,5 | 0,77 |
| Coefficient of variation of reproducibility (%) | 1,1 | 2,1 | 6,4 | 6,9 | 2,5 | 1,1 | 1,7 |
| Reproducibility limit, R (2,8 s_R) (°C) | 0,9 | 2,1 | 5,7 | 7,1 | 2,6 | 1,4 | 2,2 |

Table B.2 — Comparison of methods using palm oil samples

| Samples | | Slip point | | |
|--|---|---------------------------|----------|----------|
| | | MS 817:1989 | ISO 6321 | |
| | | AOCS Cc 3-25 ^a | Method A | Method B |
| Palm oil, refined, bleached and deodorized (RBD) | 1 | 36,8 | 38,2 | 36,5 |
| | 2 | 35,3 | 37,4 | 35,5 |
| | 3 | 35,2 | 37,7 | 35,5 |
| | 4 | 36,6 | 38,0 | 36,5 |
| | 5 | 35,6 | 37,5 | 35,5 |
| ^a Malaysian Standard MS 817:1989[3]. AOCS Official Method Cc 3-25[4]. | | | | |

2) [ISO 5725:1986](#) was used to obtain the precision data and has since been withdrawn.

| Samples | | Slip point | | |
|--|---|--|----------------------|----------|
| | | MS 817:1989 AOCS Cc 3-25 ^a | ISO 6321 Method A | Method B |
| Palm olein, RBD | 1 | 22,3 | 24,4 | 25,5 |
| | 2 | 22,2 | 24,4 | 25,5 |
| | 3 | 22,5 | 24,3 | 25,5 |
| | 4 | 22,5 | 24,2 | 24,9 |
| | 5 | 22,3 | 24,2 | 24,9 |
| Palm stearin, RBD | 1 | 51,6 | 51,8 | 51,5 |
| | 2 | 52,8 | 51,8 | 52,8 |
| | 3 | 46,0 | 44,8 | 45,0 |
| | 4 | 52,3 | 52,8 | 53,4 |
| | 5 | 51,0 | 51,0 | 51,5 |
| Crude palm oil | 1 | 35,8 | 35,6 | |
| | 2 | 35,3 | 36,6 | |
| | 3 | 35,8 | 36,4 | |
| | 4 | 35,0 | 35,8 | |
| | 5 | 35,8 | 36,8 | |
| Crude palm kernel oil | 1 | 27,8 | 27,7 | 27,6 |
| | 2 | 26,6 | 27,8 | 27,6 |
| | 3 | 26,7 | 26,7 | 27,0 |
| | 4 | 26,8 | 26,7 | 27,0 |
| | 5 | 27,0 | 27,5 | 27,4 |
| Palm kernel oil, RBD | 1 | 27,8 | 27,8 | 28,2 |
| | 2 | 27,8 | 27,6 | 27,6 |
| | 3 | 27,7 | 27,5 | 28,0 |
| | 4 | 27,8 | 27,2 | 28,0 |
| | 5 | 27,6 | 27,3 | 27,8 |
| Palm kernel olein, RBD | 1 | 26,2 | 25,8 | 26,0 |
| | 2 | 23,4 | 23,3 | 23,8 |
| | 3 | 23,5 | 23,4 | 23,8 |
| | 4 | 23,4 | 23,4 | 23,8 |
| | 5 | 24,6 | 24,4 | 24,5 |
| Palm kernel stearin, RBD | 1 | 32,2 | 32,2 | 33,0 |
| | 2 | 32,2 | 32,8 | 33,0 |
| | 3 | 39,3 | 38,5 | 39,4 |
| | 4 | 33,3 | 33,0 | 33,2 |
| | 5 | 32,3 | 33,6 | 33,2 |
| ^a Malaysian Standard MS 817:1989[3]. AOCS Official Method Cc 3-25[4]. | | | | |

Table B.3 — Statistical results for palm oil samples

| | Palm oil | Palm olein | Palm stearin |
|---|----------|------------|--------------|
| No. of laboratories retained after eliminating outliers | 10 | 11 | 11 |
| Mean (°C) | 37,4 | 20,5 | 52,1 |

| | Palm oil | Palm olein | Palm stearin |
|---|-----------------|-------------------|---------------------|
| Standard deviation of repeatability, s_r (°C) | 0,23 | 0,15 | 0,09 |
| Coefficient of variation of repeatability (%) | 0,6 | 0,7 | 0,2 |
| Repeatability limit, r (2,8 s_r)(°C) | 0,64 | 0,42 | 0,25 |
| Standard deviation of reproducibility, s_R (°C) | 0,78 | 0,98 | 0,54 |
| Coefficient of variation of reproducibility (%) | 2,1 | 4,8 | 1,0 |
| Reproducibility limit, R (2,8 s_R) (°C) | 2,2 | 2,7 | 1,5 |

Bibliography

- [1] [ISO 5555](#), *Animal and vegetable fats and oils — Sampling*
- [2] [ISO 5725:1986](#), *Precision of test methods — Determination of repeatability and reproducibility for a standard test method by inter-laboratory tests*
- [3] MS 817:1989, *The determination of melting point in open capillary tubes (slip point) for palm oil products*
- [4] AOCS Official Method Cc 3-25, *Slip melting point AOCS standard open tube melting point* (updated 1992)

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- user for personal or internal company use only.
- The standard may be stored on more than one device provided that it is accessible by the sole named user only and that only one copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than one copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright and Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email cservices@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email: cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK