



## BSI Standards Publication

### **Rubber, raw — Determination of the glass transition temperature by differential scanning calorimetry (DSC)**

---



# INTERNATIONAL STANDARD

**ISO**  
**22768**

Second edition  
2017-08

---

---

## **Rubber, raw — Determination of the glass transition temperature by differential scanning calorimetry (DSC)**

*Caoutchouc brut — Détermination de la température de transition  
vitreuse par analyse calorimétrique différentielle (DSC)*



Reference number  
ISO 22768:2017(E)

© ISO 2017



## **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
[copyright@iso.org](mailto:copyright@iso.org)  
[www.iso.org](http://www.iso.org)



# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Principle</b>	<b>1</b>
<b>5 Apparatus and materials</b>	<b>2</b>
<b>6 Test specimen</b>	<b>2</b>
<b>7 Conditioning</b>	<b>2</b>
<b>8 Calibration</b>	<b>2</b>
<b>9 Procedure</b>	<b>2</b>
9.1 Gas flow rate	2
9.2 Loading the test specimen	2
9.3 Temperature scan	3
<b>10 Expression of results</b>	<b>3</b>
<b>11 Test report</b>	<b>4</b>
<b>12 Precision</b>	<b>5</b>
<b>Annex A (informative) Precision</b>	<b>6</b>
<b>Bibliography</b>	<b>9</b>



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 45, *Rubber and rubber products*, Subcommittee SC 3, *Raw materials (including latex) for use in the rubber industry*.

This second edition cancels and replaces the first edition (ISO 22768:2006), which has been technically revised with the following changes:

- additional description on placing an empty pan (crucible) as reference;
- general DSC thermogram inserted to show an inflection point which should be  $T_g$ ;
- move of the content of the clause on precision data to an informative [Annex A](#).



# Rubber, raw — Determination of the glass transition temperature by differential scanning calorimetry (DSC)

**WARNING** — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

## 1 Scope

This document specifies a method using a differential scanning calorimeter to determine the glass transition temperature of raw rubber.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1407, *Rubber — Determination of solvent extract*

ISO 11357-1:2016, *Plastics — Differential scanning calorimetry (DSC) — Part 1: General principles*

ISO 23529, *Rubber — General procedures for preparing and conditioning test pieces for physical test methods*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11357-1 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

### 3.1

#### glass transition

reversible change in an amorphous polymer, or in amorphous regions of a partially crystalline polymer, from (or to) a rubbery or viscous condition to (or from) a glassy or hard condition

### 3.2

#### glass transition temperature

$T_g$

approximate midpoint of the temperature range over which the *glass transition* (3.1) takes place

Note 1 to entry: For the purposes of this document, the glass transition temperature is defined as the point of inflection of the DSC curve which has been obtained at a heating rate of 20 °C/min (see A.3).

## 4 Principle

The change in specific heat capacity of the rubber as a function of temperature under a specified inert atmosphere is measured using a differential scanning calorimeter (DSC). The glass transition temperature is determined from the curve thus produced.



## 5 Apparatus and materials

### 5.1 Differential scanning calorimeter, in accordance with ISO 11357-1:2016, 5.1.

The calorimeter should be operated in a room held at standard laboratory temperature. It should be protected from draughts, direct sunlight and sudden temperature changes.

### 5.2 Specimen pans (crucibles), in accordance with ISO 11357-1:2016, 5.2.

### 5.3 Gas supply, analytical grade, usually nitrogen or helium.

### 5.4 Balance, capable of measuring the specimen mass to an accuracy of $\pm 0,000\ 1$ g.

## 6 Test specimen

The test specimen shall be as representative as possible of the sample being examined and shall have a mass between 0,01 g and 0,02 g.

To determine  $T_g$  of polymers, extract raw rubber in accordance with ISO 1407.

## 7 Conditioning

Condition the sample to be examined and the test specimen in accordance with ISO 23529.

## 8 Calibration

Calibrate the calorimeter according to the manufacturer's instructions.

The use of suitable analytical grade substances is recommended to check the accuracy of the temperature scale. Ideally, substances whose melting points bracket the temperature range of interest should be chosen. *n*-Octane, *n*-heptane and cyclohexane have been found to be useful. Indium should be used if a higher temperature calibrant is required.

## 9 Procedure

### 9.1 Gas flow rate

The same inert gas flow rate with a tolerance of  $\pm 10$  %, shall be used throughout the procedure. Flow rates between 10 ml/min and 100 ml/min have been found to be suitable.

### 9.2 Loading the test specimen

Determine the mass of the test specimen to an accuracy of  $\pm 0,001$  g. The same nominal mass shall be used for all determinations. If possible, the specimen shall have a flat surface so as to give good thermal contact with the bottom of the pan.

NOTE 1 Intimate thermal contact between the test specimen and the bottom of the pan is essential for good repeatability.

Place the specimen in the pan, using tweezers and seal with a lid. Place the sealed pan in the calorimeter using tweezers.

Do not handle the test specimen or the pan with bare hands.

NOTE 2 Placing an empty pan with a lid as a reference helps to obtain stable DSC thermograms.



### 9.3 Temperature scan

**9.3.1** Cool the test specimen to a temperature of approximately  $-140\text{ }^{\circ}\text{C}$  at a rate of  $10\text{ }^{\circ}\text{C}/\text{min}$  or  $20\text{ }^{\circ}\text{C}/\text{min}$  and hold at this temperature for 1 min to 10 min until the baseline becomes stable.

A starting temperature of  $-140\text{ }^{\circ}\text{C}$  is required for the determination of rubbers with very low glass transition temperatures, e.g. high-cis polybutadiene. For rubbers with higher glass transitions, this temperature is not necessary.

A starting temperature should be chosen so that a stable base line is achieved before the glass transition region, e.g. about  $30\text{ }^{\circ}\text{C}$  to  $40\text{ }^{\circ}\text{C}$  below the expected glass transition temperature.

If the apparatus is not capable of maintaining the specified cooling rate, it should be adjusted to give a rate as close as possible to that specified.

**9.3.2** Perform the temperature scan at a heating rate of  $20\text{ }^{\circ}\text{C}/\text{min}$ , heating until a temperature about  $30\text{ }^{\circ}\text{C}$  above the upper limit of the glass transition range is reached.


NOTE Most instruments can be programmed to carry out the required thermal cycle automatically.

## 10 Expression of results

Determine the glass transition temperature as the inflection point of the transition curve using the instrument software. General DSC endothermogram and inflection point are given in [Figure 1](#).

NOTE If the glass transition temperature has to be determined directly from the curve, a better indication of the position of the inflection point is given by studying the first derivative of the curve (DDSC thermogram). In the case of the representation of the exotherm curve, this is the peak minimum.





#### Key

- A DSC thermogram
- B DDSC thermogram
- Y1 axis of DSC (exothermic direction)
- Y2 axis of DDSC
- X temperature °C
- IP inflection point

**Figure 1 — General thermogram and inflection point**

## 11 Test report

The test report shall include the following:

- a) a reference to this document, i.e. ISO 22768;
- b) identification of the sample;
- c) the mass of the specimen, in grams;
- d) the type of DSC instrument used;
- e) the type of inert gas and the flow rate;
- f) the calibrants used;
- g) the thermal cycle used;
- h) the  $T_g$  value in degrees Celsius, together with the DSC curve;
- i) the date of the test.



## **12 Precision**

See [Annex A](#).



## Annex A (informative)

### Precision

#### A.1 General

**A.1.1** The interlaboratory test programme (ITP) for precision evaluation for  $T_g$  was conducted in 2004, using the precision procedures and guidelines as described in ISO/TR 9272. Refer to ISO/TR 9272 for other details and terminology on precision evaluation.

**A.1.2** The ITP was conducted using four rubbers, NdBR (neodymium-catalysed high-*cis* BR), SBR 1502, SBR 1721 and OEESBR. These represent a range of  $T_g$  values from about  $-100\text{ }^{\circ}\text{C}$  to  $-20\text{ }^{\circ}\text{C}$ . Thirty laboratories participated in the ITP and a Type 1 precision was evaluated. A test result represents a single determination or measurement of  $T_g$  using the DSC procedure as specified in this document. Two measurements of  $T_g$  were conducted on two test days, one week apart. Measurements were conducted at both  $10\text{ }^{\circ}\text{C}/\text{min}$  and at  $20\text{ }^{\circ}\text{C}/\text{min}$  on each test day. For precision analysis, all  $T_g$  values were converted to Kelvin. This avoids negative values in the calculation algorithms and, more importantly, it provides mean values (for  $T_g$ ) that are not near zero and thus permits relative precision (in percent) to be more meaningful for all materials, i.e. it avoids large percent values.

**A.1.3** The precision results as determined by this ITP may not be applied to acceptance or rejection testing for any group of materials or products without documentation that the results of this precision evaluation actually apply to the products or materials tested.

#### A.2 Precision results

##### A.2.1 General

Precision results are given in [Table A.1](#) for each of the four materials for both  $10\text{ }^{\circ}\text{C}/\text{min}$  and  $20\text{ }^{\circ}\text{C}/\text{min}$ . The precision results were obtained using the outlier deletion procedures as described in ISO/TR 9272. [Table A.1](#) lists the number of laboratories remaining in the database for the precision evaluation after the deletion of laboratories that had outlier values. General statements for the use of the precision results are cited in [A.2.2](#). These are given in terms of both the absolute precision,  $r$  and  $R$ , and also for relative precision ( $r$ ) and ( $R$ ).

##### A.2.2 Repeatability and reproducibility statements

###### A.2.2.1 Repeatability

The repeatability, or local domain precision, for each of the materials (rubbers) has been established by the values found in [Table A.1](#), for each of the materials as listed in [Table A.1](#). Two single test results (obtained by the proper use of the method of this document) that differ by more than the tabulated values for  $r$  in measurement units, and ( $r$ ) in percent, shall be considered as suspect, i.e. to have come from different populations. Such a decision suggests that some appropriate investigative action be taken.

###### A.2.2.2 Reproducibility

The reproducibility, or global domain precision, for each of the materials has been established by the values found in [Table A.1](#), for each of the materials as listed in [Table A.1](#). Two single test results obtained in different laboratories (by the proper use of the method of this document) that differ by more than



the tabulated values for  $R$  in measurement units, and ( $R$ ) in percent, shall be considered as suspect, i.e. to have come from different populations. Such a decision suggests that some appropriate investigative action be taken.

### A.2.3 Additional analysis comments

The last column of [Table A.1](#) (the number of laboratories that were included in the database used for the final calculations for precision) indicates that a substantial number of laboratory data values were deleted as outliers. The final number of laboratories for NdBR is low because several laboratories did not submit data. There was some variation in precision improvement among the materials with outlier deletion but, on an overall basis (mean for all four materials), the repeatability limit,  $r$ , was reduced by a reduction factor ( $r_{\text{final}}/r_{\text{orig}}$ ) of 0,56 for 10 °C/min and 0,46 for 20 °C/min, after all repeatability outlier data were deleted. On the same overall basis, the reproducibility limit,  $R$ , was reduced by a reduction factor of 0,71 for both 10 °C/min and 20 °C/min procedures, after all reproducibility data outliers were deleted. Individual laboratories may show poor agreement in repeatability or poor agreement in reproducibility, or both.

There does not appear to be any substantial overall difference in the precision for 10 °C/min versus 20 °C/min. The repeatability  $r$  for 10 °C/min is 9 % greater than for 20 °C/min, while the reproducibility  $R$  for 10 °C/min is 3 % less than for 20 °C/min.

The final precision, as expressed in [Table A.1](#), represents the precision for the majority of laboratories in the ITP; these may be considered as a core group of high quality testing laboratories that constitute a benchmark level of performance for this particular property measurement.

### A.2.4 Bias

Bias is the difference between a measured average test result and a reference or true value for the measurement in question. Reference values do not exist for this test method and therefore bias cannot be evaluated.



Table A.1 — Precision data for glass transition temperature

Material	Mean level		Within laboratory			Between laboratories			Number of laboratories <sup>a</sup>
	°C	K	<i>s<sub>r</sub></i>	<i>r</i>	( <i>r</i> )	<i>s<sub>R</sub></i>	<i>R</i>	( <i>R</i> )	
Measurements at 10 °C/min									
1 — NdBR	−106,3	166,7	0,379	1,06	0,64	1,531	4,29	2,57	20
2 — SBR 1502	−54,5	218,5	0,382	1,07	0,49	1,160	3,25	1,49	25
3 — SBR 1721	−34,3	238,7	0,408	1,14	0,48	1,417	3,97	1,66	24
4 — OESSBR	−24,3	248,7	0,245	0,69	0,28	1,449	4,06	1,63	23
Pooled or average value <sup>b</sup>			0,354	0,990	0,473	1,39	3,89	1,84	
Measurements at 20 °C/min									
1 — NdBR	−104,8	168,2	0,273	0,76	0,45	1,418	3,97	2,36	19
2 — SBR 1502	−52,7	220,3	0,462	1,29	0,59	1,431	4,01	1,82	27
3 — SBR 1721	−32,2	240,8	0,372	1,04	0,43	1,164	3,26	1,35	22
4 — OESSBR	−21,9	251,1	0,190	0,53	0,21	1,724	4,83	1,92	24
Pooled or average value <sup>b</sup>			0,324	0,908	0,420	1,43	4,02	1,86	
NOTE 1 Notation used: <i>s<sub>r</sub></i> = within-laboratory standard deviation (in measurement units) <i>r</i> = repeatability (in measurement units) ( <i>r</i> ) = repeatability (percent of mean level expressed in K) <i>s<sub>R</sub></i> = between-laboratory standard deviation (for total between-laboratory variation in measurement units) <i>R</i> = reproducibility (in measurement units) ( <i>R</i> )= reproducibility (percent of mean level expressed in K) NOTE 2 The relative precision parameters, ( <i>r</i> ) and ( <i>R</i> ), are calculated using values of <i>T<sub>g</sub></i> in K. <sup>a</sup> Number of laboratories after outliers deleted; 3-step analysis; total of 30 laboratories participated. <sup>b</sup> Simple averages calculated.									

A.3 Conclusions

The glass transition temperature determined at a heating rate of 20 °C/min is about 2 °C higher than with a heating rate of 10 °C/min. The heating rate of 10 °C/min or 20 °C/min makes no significant difference to the precision. Therefore, in the interests of convenience, the faster rate has been chosen.



## Bibliography

- [1] ISO/TR 9272:2005, *Rubber and rubber products — Determination of precision for test method standards*











# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

## Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

## Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com).

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Useful Contacts

### Customer Services

**Tel:** +44 345 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 345 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)

### BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK