



BSI Standards Publication

Spices and condiments — Determination of moisture content

National foreword

This British Standard is the UK implementation of [ISO 939:2021](#).

The UK participation in its preparation was entrusted to Technical Committee AW/34, Food Technical Committee Chairmen.

A list of organizations represented on this committee can be obtained on request to its committee manager.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2021
Published by BSI Standards Limited 2021

ISBN 978 0 539 03285 7

ICS 67.220.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 January 2021.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

INTERNATIONAL STANDARD

ISO 939

Second edition
2021-01-04

Spices and condiments — Determination of moisture content

Épices et condiments — Détermination de la teneur en eau



Reference number
ISO 939:2021(E)

© ISO 2021



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO’s member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents		Page
Foreword		iv
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Principle	1
5	Reagents	1
5.1	Toluene	1
6	Apparatus	2
6.2	Reflux condenser	2
6.4	Analytical balance	2
7	Sampling	2
8	Procedure	2
8.1	Preparation of apparatus	2
8.2	Preparation of test sample	2
8.3	Test portion	2
8.4	Determination	2
9	Expression of results	3
10	Test report	3
Annex A (informative) Example of entrainment distillation apparatus		4

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 7, *Spices, culinary herbs and condiments*.

This second edition cancels and replaces the first edition (ISO 939:1980), which has been technically revised. The main changes compared with the previous edition are as follows:

- the Introduction has been deleted;
- definition [3.1](#) has been revised;
- in [8.1](#), “a potassium dichromate sulphuric acid cleaning solution” has been replaced by “a laboratory-grade detergent”;
- in [8.4](#), “asbestos cloth” has been replaced by “glass fibres, stone wool or ceramic fibres”;
- in [8.4](#), the instruction to immerse the receiver in water at room temperature has been changed from “15 min” to “45 min”;
- in [Annex A](#) b), the inner tube diameter has been changed from “9,5 to 12,5 mm” to “5 mm to 12,5 mm”.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Spices and condiments — Determination of moisture content

1 Scope

This document specifies an entrainment method for the determination of the moisture content of spices and condiments.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

[ISO 948](#), *Spices and condiments — Sampling*

[ISO 2825](#), *Spices and condiments — Preparation of a ground sample for analysis*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

moisture content

content of water extracted or collected in accordance with the method prescribed in this document

Note 1 to entry: Moisture content is expressed as a percentage by mass.

4 Principle

Determination of the amount of water entrained by azeotropic distillation, using an organic solvent immiscible with water, and collected in a graduated tube.

5 Reagents

5.1 Toluene.

Saturate the toluene by shaking with a small quantity of water and distil. Use the distillate for the determination of the moisture.

NOTE Different solvents are used for the determination of moisture.

Unless otherwise specified in the International Standard giving specifications for the spice or condiment in question, toluene shall be the solvent used for the determination.

9 Expression of results

The moisture content (w), in percentage by mass, can be expressed by [Formula \(1\)](#):

$$w = \frac{100 \times V}{m} \quad (1)$$

where

V is the volume of water collected, in ml;

m is the mass of the test portion, in g.

It is assumed that the density of water is 1 g/ml exactly.

10 Test report

The test report shall show the method used and the result obtained. It shall also mention all operating conditions not specified in this document, or regarded as optional, as well as any circumstances that may have influenced the result.

The report shall include all details required for complete identification of the sample.

Annex A (informative)

Example of entrainment distillation apparatus

The apparatus consists of a glass flask heated by a suitable means and provided with a reflux condenser discharging into a receiver connected to the flask. The connections between the receiver, the condenser and the flask are interchangeable ground glass joints. The receiver serves to collect and measure the condensed water, and to return the solvent to the flask. The assembly of the apparatus is shown in [Figure A.1](#) and the various components are described below.

- a) Flask, of capacity 500 ml, made of heat-resistant glass, well annealed, as free as possible from striae and defects, and similar to the shape shown in [Figure A.1](#).
- b) Reflux condenser, water cooled, made of glass, having a jacket approximately 400 mm long and an inner tube of diameter 5 mm to 12,5 mm. The tip of the condenser to be inserted in the receiver may be ground off at an angle of 30° from the vertical axis of the condenser. When inserted into the receiver, the tip of the condenser is 6 mm to 7 mm above the surface of the liquid in the receiver after distillation conditions have been established.
- c) Receiver, of capacity 5 ml, made of heat-resistant glass, well annealed and as free as possible from striae and similar defects, provided with ground glass joints, with the shape, dimensions and tolerances given in [Figure A.1](#), and consisting essentially of the upper chamber, together with the tube and ground joint leading to the flask, and the graduated tube. The graduated portion has a capacity of 5 ml when filled to the highest graduation mark.


The scale covers the range of 0 ml to 5 ml and is graduated at intervals of 0,1 ml. The graduation marks corresponding to each millilitre are numbered and carried completely round the tube. The graduation marks midway between the numbered marks are carried three-quarter way, and the remaining marks are carried half-way, around the tube. The error at any indicated capacity does not exceed 0,05 ml.

- d) Heat source, either an oil bath or an electric heater provided with a sliding rheostat or other means of heat control.

The temperature of the oil in the bath should not be very much higher than the boiling point of toluene.

- e) Copper wire, long enough to extend through the condenser and with one end twisted into a spiral. The diameter of the spiral is such that it fits snugly within the graduated portion of the receiver and yet can be moved up and down.

Dimensions in millimetres



Key

- | | | | |
|---|--|---|-------------------------------|
| 1 | reflux condenser (6.2) | 3 | graduated tube |
| 2 | receiver (6.3) | 4 | flask (6.1) |

NOTE Volume: 5 ml, graduations: 0,10 ml.

Figure A.1 — Assembly of the entrainment distillation apparatus

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than one device provided that it is accessible by the sole named user only and that only one copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than one copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright and Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email cservices@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email: cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

